

# REGIONAL CENTRE FOR BIOTECHNOLOGY

180, Udyog Vihar Phase-I, Gurgaon, India

Dated 01.11.2013

## CORRIGENDUM NOTICE

Please refer to the NOTICE INVITING TEDNER for Tender file No. RCB/ATN/03/13-14 for the purchase Of Plant Germination & Growth System, the undersigned is directed to incorporate the following amendments in the specifications and in terms & conditions as under:

Particulars	EXISTING	AMENDED AS
Last Date for Submission of Sealed Tenders	06.11.2013 upto 3 p.m.	20.11.2013 upto 3 p.m.
Date of opening of Technical Bids	07.11.2013 at 11.00 a.m.	21.11.2013 at 11.00 a.m.
<u>Clause No.3 of Part A of SPECIFICATIONS FOR PROGRAMMABLE PLANT GROWTH CHAMBER</u>	Internal dimension of each chamber unit should be at least 3.7ft W X 2.2ft D X 2.4ft H. External dimensions should not to exceed 6ft W X 2.4ft D X 6.5ft H.	Internal dimension of each chamber unit should be at least 3.7ft W X 2.2ft D X 2.4ft H. External dimensions should not to exceed 6ft W X 2.4ft D X 6.5ft H. <b><u>(can be relaxable to an acceptable limit)</u></b>
<u>Clause No.5 of Part B of SPECIFICATIONS FOR SEED GERMINATING CHAMBER</u>	Total unit dimensions should be at least 3.7ft W X 2.5ft D X 6.5ft H. Internal dimension of each tray should be at least 3ft X 1.7ft with at least 50mm space between trays for 20 trays.	Total unit dimensions should be at least 3.7ft W X 2.5ft D X 6.5ft H. Internal dimension of each tray should be at least 3ft X 1.7ft with at least 50mm space between trays for 20 trays. <b><u>(can be relaxable to an acceptable limit)</u></b>
<u>Clause No.19 of Part A of SPECIFICATIONS FOR PROGRAMMABLE PLANT GROWTH CHAMBER</u>	Credibility: manufacturers should be renowned international companies with a record of more than <b><u>350 numbers</u></b> of the quoted model supplied worldwide. Users list, contact numbers, and certificate (from at least 3 installations) should attached.	Credibility: manufacturers should be renowned international companies with a record of <b><u>atleast 200 numbers</u></b> of the quoted model supplied worldwide. Users list, contact numbers, and certificate (from at least 3 installations) should attached.

All the other specifications and terms & conditions of the aforementioned tender remains unaltered.

(SR. MANAGER)

For & on behalf of **EXECUTIVE DIRECTOR**