

CORRIGENDUM NOTICE

RCB/ATN/09/CORRIGENDUM/15-16/DAILAB

Dated : 19 Apr 2016

It is to intimate that for Tender No RCB/ATN/09/15-16/DAILAB for “**IN VIVO IMAGING SYSTEM**” for Regional Centre for Biotechnology, at NCR Biotech Science Cluster, Faridabad, Gurgaon-Faridabad Expressway 3rd Milestone towards Gurgaon, District Faridabad (Haryana)- 121001; following specifications has been modified.

- 1. Tender specifications point 14:** Stage movement or camera movement should be software controlled for different levels of magnifications.
- 2. Tender specifications point 15:** The system should come with an adjustable heated stage or animal chamber with warm heating within 20-40°C temperature range.

All other specifications and terms & conditions of the aforementioned tender remain unaltered.

For and on behalf of

Sr. Manager (A&F)

Executive Director, RCB